


Certificate of Analysis

Product Name: Sample Probe Cleaner
Product Numbers: 10445209/KS807
Lot Number: KP8L02

Date of Manufacture: 06-NOV-18
Expiration Date: 2019-11-06
Storage: 2 - 25 °C

Specification	Specification Limit	Result
Appearance	Clear, light yellow, free of precipitates	Complies
Normality (NaOCl)	0.54 - 0.80 N	0.69
pH at 25 °C	11.0 - 11.5	11.5

For laboratory use only.



Quality Representative

Date: 2018-11-19