

CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT PP2302
H PYLORI ISOLATION AGAR**LOT NUMBER** 1078882**EXPIRY DATE** 2009.03.25**RELEASE DATE** 2009.02.15**Physical Testing Results Specification***Tested in accordance with Oxoid Method QC13 unless otherwise stated*

Appearance (Colour)	Pass	Red
Clarity	Pass	Opaque
Gel Strength	Pass	Satisfactory
pH	7.4	7.2 - 7.6
Sterility @ 30°C for 3 days /sample plan	Pass	Within Limits

This test is performed in accordance Oxoid method QC21, NATA FADs, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1

Microbiology Testing	Result	Specification	Method
Helicobacter pylori MVCC 1231	Pass	Clear, glossy colonies	QC27
Escherichia coli ATCC® 25922	Pass	Inhibited	QC27
Candida albicans ATCC® 10231	Pass	Inhibited	QC27

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Carissa Courtney

**Approved Signatory**

Accredited for compliance with ISO/IEC 17025.
This report may not be reproduced except in full.

**MANUFACTURED AND TESTED BY:****Thermo Fisher Scientific, Microbiology Division**

20 Dalglish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054

Email: micro_orders@thermofisher.com Website: www.thermofisher.com.au