



## G & M Procter Ltd. Certificate of Analysis

**PRODUCT** EB0212E  
**COOKED MEAT MEDIUM**  
**100X10ML IN UNIVERSAL**

**LOT NUMBER** 1159614  
**EXPIRY DATE** 2012.01.20  
**PRODUCTION DATE** 2011.01.21  
**TEST DATE** 2011.01.21  
**REPORTING DATE** 2011.01.26

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Straw	Straw	Appearance and colour
pH (25°C)	7.0	6.8 - 7.2	pH
Fill Volume/Weight	10.3g	10.0 - 10.3g	Fill volume weight check
Sterility @ 22°C & 32°C ± 2°C for 5 days	No growth	No growth	Ster.at 22,32,37 & 44°C

### MICROBIOLOGICAL PERFORMANCE

For target organisms, the control media must achieve a colony count of 10-100 cfu. The test medium must show clearly visible growth as stated in the specification.

Target Organism	Control(cfu)	Test	Specification	Accredited Method Reference
Clostridium perfringens ATCC®13124	44	Turbid broth, red lighter colour meat	Turbid broth, red lighter colour meat	Fertility of Specified Target Organism(s) (Liquid Media)
Clostridium sporogenes ATCC®19404	29	Turbid broth with darkening / blackening	Turbid broth with darkening/blackening	Fertility of Specified Target Organism(s) (Liquid Media)
Staphylococcus aureus ATCC®25923	70	Turbid growth	Turbid growth	Fertility of Specified Target Organism(s) (Liquid Media)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.



M. E. McCallum

Moirra McCallum  
Quality Manager, G & M Procter Ltd.

2727

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road,  
Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727



## CERTIFICATE OF ANALYSIS

Delivery/Customer information

Date Printed

2013.11.11

Delivery No.

Customer

Customer Order number

---

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

A handwritten signature in black ink, appearing to read "M. E. McCall".

MOIRA MCCALLUM  
Quality Manager G & M Procter Ltd

Our management system is certified by BSI as being in conformity with ISO 9001:2008, certificate number FM 27644 and ISO 13485:2003, certificate number MD 85850.

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,  
4 Auld Bond Road, Perth, PH1 3FX