



Certificate of Quality

Product Name: Columbia CNA w/5% Sheep Blood 100/PK
Lot Number: 166285

Product Number: R01322
Expiration Date: 2012-05-15
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Streptococcus pyogenes	19615	Good Growth, Beta Hemolysis
Streptococcus pneumoniae	6305	Good Growth, Alpha Hemolysis
Enterococcus faecalis	29212	Good Growth
Staphylococcus aureus	25923	Good Growth
Escherichia coli	25922	No Growth
Proteus mirabilis	12453	No Growth
Pseudomonas aeruginosa	27853	No Growth

Physical Characteristics:

Appearance: Red, Opaque Agar
pH @ 20-25°C: 7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor