

Certificate of Quality

Product Name: Selective Neisseria Media QC SET/3
Lot Number: 430856

Product Number: R4651102
Expiration Date: 2019-06-28
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to performance criteria for this product.

Purity:

Standardized aliquots of the rehydrated product are inoculated onto nonselective media and examined for pure growth following the appropriate incubation. Selective and Differential media are also tested where applicable.

Viability And Quantification:

Each organism is recovered from the preserved state within the required time frame and at an acceptable level. Passage number is stated as the current preserved state.

Macroscopic And Microscopic Morphology:

Colony morphology is consistent with documented referenced description.
Traditional staining is performed.

Biochemical Analysis:

Organism exhibits characteristic biochemical and/or enzymatic reactions. Automated and/or conventional testing was performed and results were within established limits. Antimicrobial testing performed where applicable. Results within expected ranges.

	Strain	Gram Reaction:	Performance Criteria:
Neisseria gonorrhoeae	43069	Gram Negative Diplococci	Growth
Proteus mirabilis	43071	Gram Negative Rod	Partial Inhibition
Staphylococcus epidermidis	12228	Gram Positive Cocci	Partial Inhibition

CFU/1ml: >10(4)

Passage: Kit - All <5

Appearance: Preserved Gel Matrix suspended in inoculating loop
pH: N/A

Signed



Quality Assurance Supervisor